ANTAR Sp. J. I. Groniecka-Tarnkowska, A. Tarnkowski
z siedzibą w Warszawie, ul. Zawiślańska 43, 03-068 Warszawa
Tel.: +48 22 518 36 00; Fax: +48 22 518 36 30; E-mail: antar@antar.net

Warsaw, 25.05.2021

EU DECLARATION OF CONFORMITY MEDICAL DEVICE

Irena Groniecka - Tarnkowska, Andrzej Tarnkowski „ANTAR" Spółka Jawna
ul. Zawiślańska 43
03-068 Warszawa

acting as Manufacturer registered in Eudamed, SRN number: PL-MF-000001583, according to Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017, herein declare that medical devices:

ROLLATOR: AT51013, AT51014, AT51015

The Basic UDI-DI: 59015714AT510133N, 59015714AT510143Q, 59015714AT510153S

have been classifed as medical device class I, rule 1.

Intended purpose: rollators are designed for disable people to enable them to move. User rests on the rollator, makes a step forward and then pushes the rollator

We herein declare that the medical devices are in conformity with Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017. The assessment of conformity has been proceeded in compliance with the requirements of the above Regulation.

Applied standards:
PN-EN ISO 9001 (current edition)
PN-EN ISO 13485 (current edition)
PN-EN ISO 15223 (current edition)
PN-EN ISO 14971 (current edition)
PN-EN ISO 10993 (current edition)

The EU declaration of conformity is issued under the sole responsibility of the Manufacturer.

Andrzej Tarnkowski

„ANTAR" Sp. Jawna
Irena Groniecka-Tarnkowska
Andrzej Tarnkowski
ul. Zawiślańska 43, 03-068 Warszawa
NIP: 524-21-23-915, REGON 012853911
Fax: 22 518 36 30, 22 518 36 31
Tel.: 22 518 36 00

co-owner
independent reprezentation of the company
based on the Company Register

CE